



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 22-000296/D002.R000
Report Date: 01/19/2022
ORELAP#: OR100028
Purchase Order:
Received: 01/11/22 11:27

Customer: NW Natural Goods
Product identity: HEMP - RB 0069
Client/Metric ID: .
Laboratory ID: 22-000296-0001

Summary

Potency:

Analyte per 4g	Result	Limits	Units	Status	
CBD per 4g	25.3		mg/4g		CBD-Total per 4g 25.3 mg/4g
CBE per 4g [†]	0.167		mg/4g		
CBT per 4g [†]	0.620		mg/4g		THC-Total per 4g <LOQ
(Reported in milligrams per serving)					

Residual Solvents:

All analytes passing and less than LOQ.

Pesticides:

Analyte	Result (mg/kg)	Limits (mg/kg)	Status
Multi-Residue Pesticide Profile [†]	< LOQ for all analytes		

Metals:

Less than LOQ for all analytes.

Microbiology:

Less than LOQ for all analytes.



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 22-000296/D002.R000
Report Date: 01/19/2022
ORELAP#: OR100028
Purchase Order:
Received: 01/11/22 11:27

Customer: NW Natural Goods

Product identity: HEMP - RB 0069

Client/Metric ID: .

Sample Date:

Laboratory ID: 22-000296-0001

Evidence of Cooling: No

Temp: 17.5 °C

Relinquished by: Ramos

Serving Size #1: 4 g

Sample Results

Potency per 4g	Method J AOAC 2015 V98-6 (mod)	Units mg/se	Batch: 2200368	Analyze: 1/14/22 4:44:00 AM	
Analyte	Result	Limits	Units	LOQ	Notes
CBC per 4g [†]	< LOQ		mg/4g	0.128	
CBC-A per 4g [†]	< LOQ		mg/4g	0.128	
CBC-Total per 4g [†]	< LOQ		mg/4g	0.241	
CBD per 4g	25.3		mg/4g	0.128	
CBD-A per 4g	< LOQ		mg/4g	0.128	
CBD-Total per 4g	25.3		mg/4g	0.241	
CBDV per 4g [†]	< LOQ		mg/4g	0.128	
CBDV-A per 4g [†]	< LOQ		mg/4g	0.128	
CBDV-Total per 4g [†]	< LOQ		mg/4g	0.239	
CBE per 4g [†]	0.167		mg/4g	0.128	
CBG per 4g [†]	< LOQ		mg/4g	0.128	
CBG-A per 4g [†]	< LOQ		mg/4g	0.128	
CBG-Total per 4g [†]	< LOQ		mg/4g	0.239	
CBL per 4g [†]	< LOQ		mg/4g	0.128	
CBL-A per 4g [†]	< LOQ		mg/4g	0.128	
CBL-Total per 4g [†]	< LOQ		mg/4g	0.241	
CBN per 4g	< LOQ		mg/4g	0.128	
CBT per 4g [†]	0.620		mg/4g	0.128	
Δ8-THCV per 4g [†]	< LOQ		mg/4g	0.128	
Δ8-THC per 4g [†]	< LOQ		mg/4g	0.128	
Δ9-THC per 4g	< LOQ		mg/4g	0.128	
exo-THC per 4g [†]	< LOQ		mg/4g	0.128	
THC-A per 4g	< LOQ		mg/4g	0.128	
THC-Total per 4g	< LOQ		mg/4g	0.241	
THCV per 4g [†]	< LOQ		mg/4g	0.128	
THCV-A per 4g [†]	< LOQ		mg/4g	0.128	
THCV-Total per 4g [†]	< LOQ		mg/4g	0.241	
Total Cannabinoids per 4g	26.1		mg/4g		



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794

Report Number: 22-000296/D002.R000
Report Date: 01/19/2022
ORELAP#: OR100028
Purchase Order:
Received: 01/11/22 11:27



Microbiology

Analyte	Result	Limits	Units	LOQ	Batch	Analyze	Method	Status	Notes
E.coli	< LOQ		cfu/g	10	2200270	01/14/22	AOAC 991.14 (Petrifilm)		X
Total Coliforms	< LOQ		cfu/g	10	2200270	01/14/22	AOAC 991.14 (Petrifilm)		X
Mold (RAPID Petrifilm)	< LOQ		cfu/g	10	2200271	01/15/22	AOAC 2014.05 (RAPID)		X
Yeast (RAPID Petrifilm)	< LOQ		cfu/g	10	2200271	01/15/22	AOAC 2014.05 (RAPID)		X

Solvents Method Residual Solvents by GC/MS Units µg/g Batch 2200376 Analyze 01/14/22 01:28 PM

Analyte	Result	Limits	LOQ	Status	Notes	Analyte	Result	Limits	LOQ	Status	Notes
1,4-Dioxane	< LOQ	380	100	pass		2-Butanol	< LOQ	5000	200	pass	
2-Ethoxyethanol	< LOQ	160	30.0	pass		2-Methylbutane (Isopentane)	< LOQ		200		
2-Methylpentane	< LOQ		30.0			2-Propanol (IPA)	< LOQ	5000	200	pass	
2,2-Dimethylbutane	< LOQ		30.0			2,2-Dimethylpropane (neo-pentane)	< LOQ		200		
2,3-Dimethylbutane	< LOQ		30.0			3-Methylpentane	< LOQ		30.0		
Acetone	< LOQ	5000	200	pass		Acetonitrile	< LOQ	410	100	pass	
Benzene	< LOQ	2.00	1.00	pass		Butanes (sum)	< LOQ	5000	400	pass	
Cyclohexane	< LOQ	3880	200	pass		Ethyl acetate	< LOQ	5000	200	pass	
Ethyl benzene	< LOQ		200			Ethyl ether	< LOQ	5000	200	pass	
Ethylene glycol	< LOQ	620	200	pass		Ethylene oxide	< LOQ	50.0	20.0	pass	
Hexanes (sum)	< LOQ	290	150	pass		Isopropyl acetate	< LOQ	5000	200	pass	
Isopropylbenzene (Cumene)	< LOQ	70.0	30.0	pass		m,p-Xylene	< LOQ		200		
Methanol	< LOQ	3000	200	pass		Methylene chloride	< LOQ	600	60.0	pass	
Methylpropane (Isobutane)	< LOQ		200			n-Butane	< LOQ		200		
n-Heptane	< LOQ	5000	200	pass		n-Hexane	< LOQ		30.0		
n-Pentane	< LOQ		200			o-Xylene	< LOQ		200		
Pentanes (sum)	< LOQ	5000	600	pass		Propane	< LOQ	5000	200	pass	
Tetrahydrofuran	< LOQ	720	100	pass		Toluene	< LOQ	890	100	pass	
Total Xylenes	< LOQ		400			Total Xylenes and Ethyl benzene	< LOQ	2170	600	pass	

Pesticides Method AOAC 2007.01 & EN 15662 (mod) Units mg/kg Batch 2200390 Analyze 01/14/22 03:29 PM

Analyte	Result	Limits	Status	Notes
Multi-Residue Pesticide Profile†	< LOQ for all analytes			

Metals

Analyte	Result	Limits	Units	LOQ	Batch	Analyze	Method	Status	Notes
Arsenic	< LOQ	0.200	mg/kg	0.00840	2200506	01/19/22	AOAC 2013.06 (mod.)	pass	X
Cadmium	< LOQ	0.200	mg/kg	0.00840	2200506	01/19/22	AOAC 2013.06 (mod.)	pass	X
Lead	< LOQ	0.500	mg/kg	0.00840	2200506	01/19/22	AOAC 2013.06 (mod.)	pass	X
Mercury	< LOQ	0.100	mg/kg	0.00420	2200506	01/19/22	AOAC 2013.06 (mod.)	pass	X

Nutrition

Analyte	Result	Limits	Units	LOQ	Batch	Analyze	Method	Status	Notes
Moisture (Loss on Drying)	18.7		g/100g	0.10	2200369	01/13/22	AOAC 925.10 (mod.)		X
Water Activity	0.714		Aw	0.030	2200347	01/13/22	AOAC 978.18		X



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 22-000296/D002.R000
Report Date: 01/19/2022
ORELAP#: OR100028
Purchase Order:
Received: 01/11/22 11:27



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 22-000296/D002.R000
Report Date: 01/19/2022
ORELAP#: OR100028
Purchase Order:
Received: 01/11/22 11:27

These test results are representative of the individual sample selected and submitted by the client.

Abbreviations

Limits: Action Levels per OAR-333-007-0400, OAR-333-007-0210, OAR-333-007-0220, CCR title 16-division 42. BCC-section 5723

Limit(s) of Quantitation (LOQ): The minimum levels, concentrations, or quantities of a target variable (e.g., target analyte) that can be reported with a specified degree of confidence.

† = Analyte not NELAP accredited.

Units of Measure

cfu/g = Colony forming units per gram

g = Gram

g/100g = Grams per 100 Grams

µg/g = Microgram per gram

mg/kg = Milligram per kilogram = parts per million (ppm)

mg/4g = Milligram per 4g

% = Percentage of sample

Aw = Water Activity

% wt = µg/g divided by 10,000

Glossary of Qualifiers

X: Not ORELAP accredited.

Approved Signatory

Derrick Tanner
General Manager



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794

Report Number: 22-000296/D002.R000
Report Date: 01/19/2022
ORELAP#: OR100028
Purchase Order:
Received: 01/11/22 11:27



Cannabis Multi-Residue Profile, Limits of Quantitation

Compound	LOQ (mg/kg)	Compound	LOQ (mg/kg)	Compound	LOQ (mg/kg)
Abamectin	0.100	Clethodim	0.050	Endrin	0.100
Acephate	0.100	Clethodim Sulfone	0.050	EPN	0.050
Acequinocyl	0.100	Clethodim Sulfoxide	0.050	EPTC	0.100
Acetamiprid	0.020	Clofentezine	0.020	Esfenvalerate/Fenvalerate	0.200
Acetochlor	0.100	Clomazone	0.020	Etaconazole	0.100
Acrinathrin	0.100	Clothianidin	0.200	Ethalfuralin	0.100
Alachlor	0.100	Coumaphos	0.050	Ethiofencarb	0.050
Aldicarb	0.100	Crotoxyphos	0.020	Ethion	0.200
Aldicarb sulfoxide	0.100	Cyanazine	0.020	Ethirimol	0.100
Aldoxycarb (Aldicarb-sulfone)	0.100	Cyanofenphos	0.020	Ethofumesate	0.050
Aldrin	0.100	Cyantranilprole	0.050	Ethoprophos	0.020
Ametoctradin	0.020	Cyazofamid	0.020	Etofenprox	0.020
Ametryn	0.500	Cycloate	0.100	Etoxazole	0.020
Aspon	0.100	Cyfluthrin	0.200	Etridiazole	0.100
Asulam	0.100	Cyhalothrin, lambda	0.200	Etrimfos	0.020
Atrazine	0.100	Cymoxanil	0.050	Famoxadone	0.200
Atrazine-desethyl	0.100	Cypermethrin	0.200	Famphur	0.100
Azinphos-ethyl	0.020	Cyprodinil	0.100	Fenamidone	0.020
Azinphos-methyl	0.020	Dacthal	0.100	Fenamiphos	0.020
Azoxystrobin	0.020	Daminozide	0.100	Fenamiphos sulfone	0.020
Benalaxyl	0.020	DCPMU	0.050	Fenamiphos sulfoxide	0.020
Bendiocarb	0.020	DDD, o,p'-	0.100	Fenazaquin	0.100
Benfluralin	0.100	DDD, p,p'-	0.100	Fenbuconazole	0.100
Benoxacor	0.050	DDE, o,p'-	0.100	Fenclorphos	0.100
Bensulide	0.050	DDE, p,p'-	0.100	Fenclorphos-oxon	0.100
BHC alpha isomer	0.100	DDT, o,p'-	0.100	Fenhexamid	0.100
BHC beta isomer	0.100	DDT, p,p'-	0.100	Fenitrothion	0.100
BHC delta isomer	0.500	DEF (Tribufos)	0.100	Fenobucarb	0.050
Bifenazate	0.020	Deltamethrin	0.100	Fenoxycarb	0.020
Bifenthrin	0.020	Desmedipham	0.100	Fenpropathrin	0.050
Boscalid	0.020	Diallate	0.100	Fenpyroximate	0.020
Bromophos-ethyl	0.100	Diazinon	0.020	Fenson	0.100
Bromophos-methyl	0.200	Diazoxon	0.100	Fensulfthion	0.020
Bromopropylate	0.100	Dichlobenil	0.100	Fensulfthion oxon	0.020
Bromuconazole	0.100	Dichlofluanid	0.100	Fensulfthion sulfone	0.100
Bupirimate	0.020	Dichlorvos	0.100	Fensulfthion-oxon-sulfone	0.020
Buprofezin	0.050	Diclobutrazol	0.050	Fenthion	0.050
Butachlor	0.500	Dicofol	0.100	Fenthion oxon	0.020
Butralin	0.200	Dicrotophos	0.050	Fenthion oxon sulfone	0.100
Butylate	0.100	Dieldrin	0.100	Fenthion sulfone	0.050
Cadusafos	0.020	Diethofencarb	0.020	Fenuron	0.020
Captan	1.000	Diethyltoluamide (DEET)	0.050	Fipronil	0.100
Carbaryl	0.050	Difenoconazole	0.100	Fonicamid	0.100
Carbendazim	0.100	Dimethenamid	0.050	Fluchloralin	0.100
Carbofuran	0.020	Dimethoate	0.050	Flucythrinate	0.100
Carbophenothion	0.200	Dimethomorph	0.050	Fludioxonil	0.200
Carboxin	0.020	Diniconazole	0.200	Flufenacet	0.020
Carfentrazone-ethyl	0.100	Dinotefuran	0.200	Flumioxazin	0.100
Chlorantranilprole	0.020	Dioxathion	0.100	Fluometuron	0.020
Chlordane, cis-	0.200	Diphenamid	0.020	Fluopicolide	0.050
Chlordane, trans-	0.200	Diphenylamine	0.100	Fluopyram	0.020
Chlorfenapyr	0.500	Disulfoton	0.100	Fluoxastrobin	0.050
Chlorfenson	0.200	Disulfoton sulfone	0.100	Flupyradifurone	0.020
Chlorfenvinphos	0.050	Disulfoton sulfoxide	0.100	Fluridone	0.100
Chlorobenzilate	0.100	Diuron	0.050	Flusilazole	0.020
Chloroneb	0.200	Edifenphos	0.050	Flutolanil	0.020
Chlorpyrifos	0.050	Endosulfan alpha	0.200	Flutriafol	0.020
Chlorpyrifos-methyl	0.200	Endosulfan beta	0.200	Fluvalinate, tau-	0.100
CIPC	1.000	Endosulfan sulfate	0.100	Fluxapyroxad	0.020



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794

Report Number: 22-000296/D002.R000
Report Date: 01/19/2022
ORELAP#: OR100028
Purchase Order:
Received: 01/11/22 11:27



Cannabis Multi-Residue Profile, Limits of Quantitation

Compound	LOQ (mg/kg)	Compound	LOQ (mg/kg)	Compound	LOQ (mg/kg)
Fomesafen	0.100	Mexacarbate	0.020	Propamocarb	0.050
Fonofos	0.100	MGK 264	0.020	Propanil	0.050
Forchlorfenuron	0.050	Mirex	0.100	Propargite	0.050
Formetanate	0.050	Molinate	0.050	Propazine	0.020
Furathiocarb	0.020	Monocrotophos	0.100	Propetamphos	0.050
Heptachlor	0.100	Monolinuron	0.020	Propham	0.050
Heptachlor epoxide	0.100	Myclobutanil	0.050	Propiconazole	0.050
Heptenophos	0.100	Naled	0.100	Propoxur	0.050
Hexachlorobenzene	0.100	Napropamide	0.050	Propoxycarbazone Na	0.050
Hexaconazole	0.100	Neburon	0.020	Propyzamide	0.050
Hexazinone	0.100	Nitrapyrin	0.100	Prothiofos	0.100
Hexythiazox	0.020	Norflurazon	0.050	Pyraclostrobin	0.020
Imazalil	0.100	Omethoate	0.100	Pyrazophos	0.050
Imidacloprid	0.100	O-Phenylphenol	0.100	Pyrethrins	0.050
Indaziflam	0.020	Oxadixyl	0.100	Pyridaben	0.020
Indoxacarb	0.020	Oxamyl	0.100	Pyridafol	0.100
Iprobenfos	0.100	Oxamyl-oxime	0.100	Pyridate	0.020
Iprodione	0.100	Oxychlordan	0.100	Pyrimethanil	0.050
Isobenzan	0.100	Oxydemeton-Methyl	0.100	Pyriproxifen	0.020
Isocarbophos	0.500	Oxythioquinox	0.200	Pyroxasulfone	0.020
Isodrin	0.100	Pacllobutrazol	0.050	Pyroxsulam	0.020
Isfenphos	0.050	Paraoxon-ethyl	0.020	Quinalphos	0.050
Isfenphos-methyl	0.020	Paraoxon-methyl	0.100	Quinoxyfen	0.050
Isfenphos oxon	0.050	Parathion ethyl	0.100	Quintozene (PCNB)	0.200
Isoprocarb	0.020	Parathion methyl	0.200	Resmethrin	0.050
Isopropalin	0.200	Penconazole	0.050	Rotenone	0.050
Isoprothiolane	0.050	Pendimethalin	0.050	S421	0.100
Isoproturon	0.050	Penflufen	0.020	Simazine	0.100
Isoxaben	0.050	Pentachloroaniline	0.100	Simetryn	0.200
Isoxaflutole	0.050	Pentachloroanisole	0.100	Spinetoram	0.020
Kresoxim-methyl	0.050	Pentachlorobenzene (PCB)	0.100	Spinosad	0.050
Lactofen	0.500	Pentachlorothioanisole (PCTA)	0.100	Spirodiclofen	0.100
Lenacil	0.100	Penthiopyrad	0.020	Spiromesifen	0.050
Lindane (gamma BHC)	0.100	Permethrin	0.050	Spirotetramat	0.050
Linuron	0.020	Perthane	0.100	Spiroxamine	0.020
Malaaxon	0.050	Phenmedipharm	0.050	Sulfotep	0.050
Malathion	0.050	Phenthoate	0.050	Sulfoxaflor	0.050
Mandipropamid	0.020	Phorate	0.050	Sulprofos	0.020
Mecarbam	0.020	Phorate Sulfone	0.050	Tebuconazole	0.100
Mepanipyrim	0.050	Phorate Sulfoxide	0.050	Tebufenozide	0.020
Merphos	0.500	Phosalone	0.050	Tebuthiuron	0.020
Metalaxyl	0.050	Phosmet	0.100	Tecnazene	0.100
Metaldehyde	0.050	Phosphamidon	0.050	Tefluthrin	0.100
Metconazole	0.100	Phoxim	0.050	Terbufos	0.020
Methacrifos	0.100	Pinoxaden	0.020	Terbufos sulfone	0.050
Methamidophos	0.050	Piperonyl butoxide	0.050	Terbufos sulfoxide	0.050
Methidathion	0.050	Pirimicarb	0.020	Terbuthylazine	0.020
Methiocarb	0.050	Pirimiphos-methyl	0.050	Terbutryn	0.020
Methiocarb sulfone	0.100	Pirimiphos-ethyl	0.020	Tetrachlorvinphos	0.050
Methiocarb sulfoxide	0.100	Prallethrin	0.100	Tetraconazole	0.050
Methomyl	0.100	Prochloraz	0.020	Tetradifon	0.200
Methoxychlor	0.100	Procymidone	0.100	Tetramethrin	0.050
Methoxyfenozide	0.020	Profenofos	0.100	Tetrasul	0.100
Metobromuron	0.050	Profluralin	0.100	Thiabendazole	0.100
Metolachlor	0.100	Promecarb	0.050	Thiabendazole, 5-hydroxy	0.100
Metolcarb	0.050	Prometon	0.100	Thiacloprid	0.050
Metrafenone	0.050	Prometryn	0.020	Thiamethoxam	0.100
Metribuzin	0.100	Propachlor	0.020	Thiobencarb	0.050
Mevinphos	0.100			Thiodicarb	0.050
				Thiophanate-methyl	0.050



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 22-000296/D002.R000
Report Date: 01/19/2022
ORELAP#: OR100028
Purchase Order:
Received: 01/11/22 11:27



Cannabis Multi-Residue Profile, Limits of Quantitation

Compound	LOQ (mg/kg)	Compound	LOQ (mg/kg)	Compound	LOQ (mg/kg)
Tolclofos-methyl	0.100	Triazophos	0.020	Trifloxystrobin	0.020
Triforin	0.100	Tolyfluanid	0.050	Triticonazole	0.050
Tralkoxydim	0.100	Tridiphane	0.500	Vinclozolin	0.100
Triadimefon	0.050	Triflumizole	0.020	Zoxamide	0.020
Triallate	0.100	Trifluralin	0.100		

LOQ = Limit of Quantitation, mg/kg

Factors affecting the LOQ include instrumentation sensitivity for a particular analyte, sample size, moisture content (percent solids) of the sample, effectiveness of the cleanup on the sample extract, and especially the type of sample matrix.



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 22-000296/D002.R000
Report Date: 01/19/2022
ORELAP#: OR100028
Purchase Order:
Received: 01/11/22 11:27



12423 NE Whitaker Way Portland OR 97230 p.503-254-1794

Cannabis Chain of Custody Record

ORELAP ID: OR100028

Company: NW Natural Goods		Analysis Requested												Purchase Order Number:																										
Contact: Isaac Velasquez		<table border="1"> <tr> <td colspan="12">OPEN MARKET</td> </tr> <tr> <td colspan="12" style="text-align: center;">CC PU</td> </tr> </table>												OPEN MARKET												CC PU												Project Number:		
OPEN MARKET																																								
CC PU																																								
Address: 11791 SE HWY 212, Clackamas, OR 97015														Project Name:																										
Email: isaacv@nwnaturalgoods.com		<input type="checkbox"/> Report Instructions: <input type="checkbox"/> Send to State - METRC <input checked="" type="checkbox"/> Email Final Results: <input type="checkbox"/> Fax Final Results <input type="checkbox"/> Cash/Check/CC/Net 30																																						
Phone: 818-644-9479 Fax:		Other:																																						
Processor's License: 330-1058115IHH																																								
Field ID	Date/Time Collected	Pesticides - OR 59 compounds	Pesticide Multi-Residue - 379 compounds	Potency	Residual Solvents	Water Activity	Moisture	Terpenes	Micro: Yeast and Mold	Micro: E Coli and Total Coliform	Heavy Metals	Mycotoxins	Other	Matrix	Weight	Serving size for edibles	Comments/Metric ID																							
HEMP - RB 0069	1/11/22	X	X	X	X	X	X	X	X	X	X			edible	40g	4g																								

Collected By:	Relinquished By:	Date	Time	Received by:	Date	Time	Lab Use Only:
<input checked="" type="checkbox"/> Standard (5 day)	Isaac Velasquez	1/11/22	1005	<i>[Signature]</i>	1.11.22	1005	Client Alias:
<input type="checkbox"/> Rush (3-4 day) (1.5x Standard)	<i>[Signature]</i>	1.11.22	1113	<i>[Signature]</i>	1/11/22	1127	Order Number:
<input type="checkbox"/> Priority Rush (2 day) (2x Standard)							Proper Container
							Sample Condition
							Temperature: 17.5
							Shipped Via:
							Evidence of cooling: <input type="checkbox"/> Yes <input type="checkbox"/> No

SUBMISSION OF SAMPLES WITH TESTING REQUIREMENTS TO PIXIS WILL BE UNDERSTOOD TO BE AN AGREEMENT FOR SERVICES IN ACCORDANCE WITH THE CONDITIONS LISTED ON THE BACK OF THIS FORM
 Revision: 1.00 Control#: CF023 www.pixislabs.com
 Effective 11/8/2018 Revised 11/8/2018 Page 1 of 2



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794

Report Number: 22-000296/D002.R000
Report Date: 01/19/2022
ORELAP#: OR100028
Purchase Order:
Received: 01/11/22 11:27



Revision #: 0.00 Control : CFL-D06
Revision Date: 05/31/2019 Effective Date: 05/31/2019

Laboratory Quality Control Results

JAOAC2015 V98-6							
Batch ID: 2200368							
Laboratory Control Sample							
Analyte	Result	Spike	Units	% Rec	Limits	Evaluation	Notes
CBDVA	0.00992	0.01	%	99.2	85.0 - 115	Acceptable	
CBDV	0.00999	0.01	%	99.9	85.0 - 115	Acceptable	
CBE	0.0103	0.01	%	103	85.0 - 115	Acceptable	
CBD A	0.0106	0.01	%	106	85.0 - 115	Acceptable	
CBGA	0.00992	0.01	%	99.2	85.0 - 115	Acceptable	
CBG	0.00994	0.01	%	99.4	85.0 - 115	Acceptable	
CBD	0.0106	0.01	%	106	85.0 - 115	Acceptable	
THCV	0.00970	0.01	%	97.0	85.0 - 115	Acceptable	
d8THCV	0.0102	0.01	%	102	85.0 - 115	Acceptable	
THCVA	0.00954	0.01	%	95.4	85.0 - 115	Acceptable	
CBN	0.00999	0.01	%	99.9	85.0 - 115	Acceptable	
exo-THC	0.00958	0.01	%	95.8	85.0 - 115	Acceptable	
d9THC	0.00980	0.01	%	98.0	85.0 - 115	Acceptable	
d8THC	0.00880	0.01	%	88.0	85.0 - 115	Acceptable	
CBL	0.00927	0.01	%	92.7	85.0 - 115	Acceptable	
CB C	0.00898	0.01	%	89.8	85.0 - 115	Acceptable	
THCA	0.0105	0.01	%	105	85.0 - 115	Acceptable	
CBCA	0.00978	0.01	%	97.8	85.0 - 115	Acceptable	
CBLA	0.00992	0.01	%	99.2	85.0 - 115	Acceptable	
CBT	0.00866	0.01	%	86.6	85.0 - 115	Acceptable	

Method Blank

Analyte	Result	LOQ	Units	Limits	Evaluation	Notes
CBDVA	<LOQ	0.0006	%	< 0.0006	Acceptable	
CBDV	<LOQ	0.0006	%	< 0.0006	Acceptable	
CBE	<LOQ	0.0006	%	< 0.0006	Acceptable	
CBD A	<LOQ	0.0006	%	< 0.0006	Acceptable	
CBGA	<LOQ	0.0006	%	< 0.0006	Acceptable	
CBG	<LOQ	0.0006	%	< 0.0006	Acceptable	
CBD	<LOQ	0.0006	%	< 0.0006	Acceptable	
THCV	<LOQ	0.0006	%	< 0.0006	Acceptable	
d8THCV	<LOQ	0.0006	%	< 0.0006	Acceptable	
THCVA	<LOQ	0.0006	%	< 0.0006	Acceptable	
CBN	<LOQ	0.0006	%	< 0.0006	Acceptable	
exo-THC	<LOQ	0.0006	%	< 0.0006	Acceptable	
d9THC	<LOQ	0.0006	%	< 0.0006	Acceptable	
d8THC	<LOQ	0.0006	%	< 0.0006	Acceptable	
CBL	<LOQ	0.0006	%	< 0.0006	Acceptable	
CB C	<LOQ	0.0006	%	< 0.0006	Acceptable	
THCA	<LOQ	0.0006	%	< 0.0006	Acceptable	
CBCA	<LOQ	0.0006	%	< 0.0006	Acceptable	
CBLA	<LOQ	0.0006	%	< 0.0006	Acceptable	
CBT	<LOQ	0.0006	%	< 0.0006	Acceptable	

Abbreviations

ND - None Detected at or above MRL
RPD - Relative Percent Difference
LOQ - Limit of Quantitation

Units of Measure:

%- Percent



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 22-000296/D002.R000
Report Date: 01/19/2022
ORELAP#: OR100028
Purchase Order:
Received: 01/11/22 11:27

Revision #: 0.00 Control : CFL-D06
Revision Date: 05/31/2019 Effective Date: 05/31/2019

Laboratory Quality Control Results

JAOAC2015 V98-6		Batch ID: 2200368							
Sample Duplicate		Sample ID: 22-000100-0001-01							
Analyte	Result	Org. Result	LOQ	Units	RPD	Limits	Evaluation	Notes	
CBDVA	<LOQ	<LOQ	0.0006	%	NA	< 20	Acceptable		
CBDV	<LOQ	<LOQ	0.0006	%	NA	< 20	Acceptable		
CBE	<LOQ	<LOQ	0.0006	%	NA	< 20	Acceptable		
CBD A	<LOQ	<LOQ	0.0006	%	NA	< 20	Acceptable		
CBG A	<LOQ	<LOQ	0.0006	%	NA	< 20	Acceptable		
CBG	<LOQ	<LOQ	0.0006	%	NA	< 20	Acceptable		
CBD	0.117	0.118	0.0006	%	0.600	< 20	Acceptable		
THCV	<LOQ	<LOQ	0.0006	%	NA	< 20	Acceptable		
d8THCV	<LOQ	<LOQ	0.0006	%	NA	< 20	Acceptable		
THCVA	<LOQ	<LOQ	0.0006	%	NA	< 20	Acceptable		
CBN	<LOQ	<LOQ	0.0006	%	NA	< 20	Acceptable		
exo-THC	<LOQ	<LOQ	0.0006	%	NA	< 20	Acceptable		
d9THC	<LOQ	<LOQ	0.0006	%	NA	< 20	Acceptable		
d8THC	<LOQ	<LOQ	0.0006	%	NA	< 20	Acceptable		
CBL	<LOQ	<LOQ	0.0006	%	NA	< 20	Acceptable		
CBC	<LOQ	<LOQ	0.0006	%	NA	< 20	Acceptable		
THCA	<LOQ	<LOQ	0.0006	%	NA	< 20	Acceptable		
CBCA	<LOQ	<LOQ	0.0006	%	NA	< 20	Acceptable		
CBLA	<LOQ	<LOQ	0.0006	%	NA	< 20	Acceptable		
CBT	<LOQ	<LOQ	0.0006	%	NA	< 20	Acceptable		

Abbreviations

- ND - None Detected at or above MRL
- RPD - Relative Percent Difference
- LOQ - Limit of Quantitation
- NA - Calculation Not Applicable given non-numerical results

Units of Measure:

%- Percent



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 22-000296/D002.R000
Report Date: 01/19/2022
ORELAP#: OR100028
Purchase Order:
Received: 01/11/22 11:27

Revision: Document ID:
Legacy ID: Effective:

Laboratory Quality Control Results									
Residual Solvents					Batch ID: 2200376				
Method Blank			Laboratory Control Sample						
Analyte	Result	LOQ	Notes	Result	Spike	Units	% Rec	Limits	Notes
Propane	ND	< 200		397	401	µg/g	99.0	70 - 130	
Isobutane	ND	< 200		514	498	µg/g	103.2	70 - 130	
Butane	ND	< 200		512	493	µg/g	103.9	70 - 130	
2,2-Dimethylpropane	ND	< 200		574	628	µg/g	91.4	70 - 130	
Methanol	ND	< 200		1610	1610	µg/g	100.0	70 - 130	
Ethylene Oxide	ND	< 30		37.4	37.2	µg/g	100.5	70 - 130	
2-Methylbutane	ND	< 200		1680	1630	µg/g	103.1	70 - 130	
Pentane	ND	< 200		1690	1610	µg/g	105.0	70 - 130	
Ethanol	ND	< 200		1850	1630	µg/g	113.5	70 - 130	
Ethyl Ether	ND	< 200		1710	1610	µg/g	106.2	70 - 130	
2,2-Dimethylbutane	ND	< 30		187	165	µg/g	113.3	70 - 130	
Acetone	ND	< 200		1710	1610	µg/g	106.2	70 - 130	
2-Propanol	ND	< 200		1820	1610	µg/g	113.0	70 - 130	
Ethyl Formate	ND	< 500		1660	1620	µg/g	102.5	70 - 130	
Acetonitrile	ND	< 100		606	498	µg/g	121.7	70 - 130	
Methyl Acetate	ND	< 500		1930	1810	µg/g	106.6	70 - 130	
2,3-Dimethylbutane	ND	< 30		172	162	µg/g	106.2	70 - 130	
Dichloromethane	ND	< 60		526	498	µg/g	105.6	70 - 130	
2-Methylpentane	ND	< 30		184	167	µg/g	110.2	70 - 130	
MTBE	ND	< 500		1650	1610	µg/g	102.5	70 - 130	
3-Methylpentane	ND	< 30		193	179	µg/g	107.8	70 - 130	
Hexane	ND	< 30		172	164	µg/g	104.9	70 - 130	
1-Propanol	ND	< 500		1620	1620	µg/g	100.0	70 - 130	
Methylethylketone	ND	< 500		1700	1770	µg/g	96.0	70 - 130	
Ethyl acetate	ND	< 200		1730	1620	µg/g	106.8	70 - 130	
2-Butanol	ND	< 200		2010	1600	µg/g	125.6	70 - 130	
Tetrahydrofuran	ND	< 100		557	500	µg/g	111.4	70 - 130	
Cyclohexane	ND	< 200		1720	1610	µg/g	106.8	70 - 130	
2-methyl-1-propanol	ND	< 500		1530	1610	µg/g	95.0	70 - 130	
Benzene	ND	< 1		6.52	5.62	µg/g	116.0	70 - 130	
Isopropyl Acetate	ND	< 200		1770	1610	µg/g	109.9	70 - 130	
Heptane	ND	< 200		1760	1610	µg/g	109.3	70 - 130	
1-Butanol	ND	< 500		1720	1620	µg/g	106.2	70 - 130	
Propyl Acetate	ND	< 500		1760	1620	µg/g	108.6	70 - 130	
1,4-Dioxane	ND	< 100		579	502	µg/g	115.3	70 - 130	
2-Ethoxyethanol	ND	< 30		180	164	µg/g	109.8	70 - 130	
Methylisobutylketone	ND	< 500		1700	1620	µg/g	104.9	70 - 130	
3-Methyl-1-butanol	ND	< 500		1540	1620	µg/g	95.1	70 - 130	
Ethylene Glycol	ND	< 200		736	502	µg/g	146.6	70 - 130	Q1
Toluene	ND	< 200		562	488	µg/g	115.2	70 - 130	
Isobutyl Acetate	ND	< 500		1860	1700	µg/g	109.4	70 - 130	
1-Pentanol	ND	< 500		1630	1630	µg/g	100.0	70 - 130	
Butyl Acetate	ND	< 500		1950	1660	µg/g	117.5	70 - 130	
Ethylbenzene	ND	< 200		1220	965	µg/g	126.4	70 - 130	
m,p-Xylene	ND	< 200		1260	990	µg/g	127.3	70 - 130	
o-Xylene	ND	< 200		1230	971	µg/g	126.7	70 - 130	
Cumene	ND	< 30		230	179	µg/g	128.5	70 - 130	
Anisole	ND	< 500		1510	1650	µg/g	91.5	70 - 130	
DMSO	ND	< 500		1750	1630	µg/g	107.4	70 - 130	
1,2-dimethoxyethane	ND	< 50		202	183	µg/g	110.4	70 - 130	
Triethylamine	ND	< 500		1660	1620	µg/g	102.5	70 - 130	
N,N-dimethylformamide	ND	< 150		488	495	µg/g	98.6	70 - 130	
N,N-dimethylacetamide	ND	< 150		629	502	µg/g	125.3	70 - 130	
Pyridine	ND	< 50		183	186	µg/g	98.4	70 - 130	
1,2-Dichloroethane	ND	< 1		1.25	1	µg/g	125.0	70 - 130	
Chloroform	ND	< 1		1.31	1	µg/g	131.0	70 - 130	Q1
Trichloroethylene	ND	< 1		1.23	1	µg/g	123.0	70 - 130	



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 22-000296/D002.R000
Report Date: 01/19/2022
ORELAP#: OR100028
Purchase Order:
Received: 01/11/22 11:27

Revision: Document ID:
Legacy ID: Effective:

QC - Sample Duplicate Sample ID: 22-000320-0001

Analyte	Result	Org. Result	LOQ	Units	RPD	Limits	Accept/Fail	Notes
Propane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Isobutane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Butane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2,2-Dimethylpropane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Methanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethylene Oxide	ND	ND	30	µg/g	0.0	< 20	Acceptable	
2-Methylbutane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Pentane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethyl Ether	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2,2-Dimethylbutane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Acetone	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2-Propanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethyl Formate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Acetonitrile	ND	ND	100	µg/g	0.0	< 20	Acceptable	
Methyl Acetate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
2,3-Dimethylbutane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Dichloromethane	ND	ND	60	µg/g	0.0	< 20	Acceptable	
2-Methylpentane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
MTBE	ND	ND	500	µg/g	0.0	< 20	Acceptable	
3-Methylpentane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Hexane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
1-Propanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Methylcyclohexane	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Ethyl acetate	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2-Butanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Tetrahydrofuran	ND	ND	100	µg/g	0.0	< 20	Acceptable	
Cyclohexane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2-methyl-1-propanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Benzene	ND	ND	1	µg/g	0.0	< 20	Acceptable	
Isopropyl Acetate	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Heptane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
1-Butanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Propyl Acetate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
1,4-Dioxane	ND	ND	100	µg/g	0.0	< 20	Acceptable	
2-Ethoxyethanol	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Methylisobutylketone	ND	ND	500	µg/g	0.0	< 20	Acceptable	
3-Methyl-1-butanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Ethylene Glycol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Toluene	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Isobutyl Acetate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
1-Pentanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Butyl Acetate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Ethylbenzene	ND	ND	200	µg/g	0.0	< 20	Acceptable	
m,p-Xylene	ND	ND	200	µg/g	0.0	< 20	Acceptable	
o-Xylene	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Cumene	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Anisole	ND	ND	500	µg/g	0.0	< 20	Acceptable	
DMSO	ND	ND	500	µg/g	0.0	< 20	Acceptable	
1,2-dimethoxyethane	ND	ND	50	µg/g	0.0	< 20	Acceptable	
Triethylamine	ND	ND	500	µg/g	0.0	< 20	Acceptable	
N,N-dimethylformamide	ND	ND	150	µg/g	0.0	< 20	Acceptable	
N,N-dimethylacetamide	ND	ND	150	µg/g	0.0	< 20	Acceptable	
Pyridine	ND	ND	50	µg/g	0.0	< 20	Acceptable	
1,2-Dichloroethane	ND	ND	1	µg/g	0.0	< 20	Acceptable	
Chloroform	ND	ND	1	µg/g	0.0	< 20	Acceptable	
Trichloroethylene	ND	ND	1	µg/g	0.0	< 20	Acceptable	

Abbreviations

ND - None Detected at or above MRL
RPD - Relative Percent Difference
LOQ - Limit of Quantitation

Units of Measure:

µg/g - Microgram per gram or ppm



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 22-000296/D002.R000
Report Date: 01/19/2022
ORELAP#: OR100028
Purchase Order:
Received: 01/11/22 11:27

Explanation of QC Flag Comments:

Code	Explanation
Q	Matrix interferences affecting spike or surrogate recoveries.
Q1	Quality control result biased high. Only non-detect samples reported.
Q2	Quality control outside QC limits. Data considered estimate.
Q3	Sample concentration greater than four times the amount spiked.
Q4	Non-homogenous sample matrix, affecting RPD result and/or % recoveries.
Q5	Spike results above calibration curve.
Q6	Quality control outside QC limits. Data acceptable based on remaining QC.
R	Relative percent difference (RPD) outside control limit.
R1	RPD non-calculable, as sample or duplicate results are less than five times the LOQ.
R2	Sample replicates RPD non-calculable, as only one replicate is within the analytical range.
LOQ1	Quantitation level raised due to low sample volume and/or dilution.
LOQ2	Quantitation level raised due to matrix interference.
B	Analyte detected in method blank, but not in associated samples.
B1	The sample concentration is greater than 5 times the blank concentration.
B2	The sample concentration is less than 5 times the blank concentration.



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 22-003713/D002.R000
Report Date: 04/07/2022
ORELAP#: OR100028
Purchase Order:
Received: 04/01/22 11:10

Customer: NW Natural Goods
Product identity: HEMP - BB 0071
Client/Metric ID: .
Laboratory ID: 22-003713-0001

Summary

Potency:

Analyte per 4g	Result	Limits	Units	Status	
CBC per 4g [†]	0.266		mg/4g		CBD-Total per 4g 26.8 mg/4g
CBD per 4g	26.8		mg/4g		
CBN per 4g	0.146		mg/4g		THC-Total per 4g <LOQ
CBT per 4g [†]	0.198		mg/4g		(Reported in milligrams per serving)

Residual Solvents:

All analytes passing and less than LOQ.

Pesticides:

Analyte	Result (mg/kg)	Limits (mg/kg)	Status
Multi-Residue Pesticide Profile [†]	< LOQ for all analytes		

Metals:

Less than LOQ for all analytes.

Microbiology:

Less than LOQ for all analytes.



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 22-003713/D002.R000
Report Date: 04/07/2022
ORELAP#: OR100028
Purchase Order:
Received: 04/01/22 11:10

Customer: NW Natural Goods

Product identity: HEMP - BB 0071

Client/Metric ID: .

Sample Date:

Laboratory ID: 22-003713-0001

Evidence of Cooling: No

Temp: 21.6 °C

Relinquished by: Ramos

Serving Size #1: 4 g

Sample Results

Potency per 4g					
Method J AOAC 2015 V98-6 (mod) Units mg/se Batch: 2202906 Analyze: 4/5/22 4:21:00 AM					
Analyte	Result	Limits	Units	LOQ	Notes
CBC per 4g [†]	0.266		mg/4g	0.131	
CBC-A per 4g [†]	< LOQ		mg/4g	0.131	
CBC-Total per 4g [†]	0.266		mg/4g	0.247	
CBD per 4g	26.8		mg/4g	0.131	
CBD-A per 4g	< LOQ		mg/4g	0.131	
CBD-Total per 4g	26.8		mg/4g	0.247	
CBDV per 4g [†]	< LOQ		mg/4g	0.131	
CBDV-A per 4g [†]	< LOQ		mg/4g	0.131	
CBDV-Total per 4g [†]	< LOQ		mg/4g	0.245	
CBE per 4g [†]	< LOQ		mg/4g	0.131	
CBG per 4g [†]	< LOQ		mg/4g	0.131	
CBG-A per 4g [†]	< LOQ		mg/4g	0.131	
CBG-Total per 4g [†]	< LOQ		mg/4g	0.245	
CBL per 4g [†]	< LOQ		mg/4g	0.131	
CBL-A per 4g [†]	< LOQ		mg/4g	0.131	
CBL-Total per 4g [†]	< LOQ		mg/4g	0.247	
CBN per 4g	0.146		mg/4g	0.131	
CBT per 4g [†]	0.198		mg/4g	0.131	
Δ8-THCV per 4g [†]	< LOQ		mg/4g	0.131	
Δ8-THC per 4g [†]	< LOQ		mg/4g	0.131	
Δ9-THC per 4g	< LOQ		mg/4g	0.131	
exo-THC per 4g [†]	< LOQ		mg/4g	0.131	
THC-A per 4g	< LOQ		mg/4g	0.131	
THC-Total per 4g	< LOQ		mg/4g	0.247	
THCV per 4g [†]	< LOQ		mg/4g	0.131	
THCV-A per 4g [†]	< LOQ		mg/4g	0.131	
THCV-Total per 4g [†]	< LOQ		mg/4g	0.247	
Total Cannabinoids per 4g	27.4		mg/4g		



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 22-003713/D002.R000
Report Date: 04/07/2022
ORELAP#: OR100028
Purchase Order:
Received: 04/01/22 11:10

Microbiology

Analyte	Result	Limits	Units	LOQ	Batch	Analyze	Method	Status	Notes
E.coli	< LOQ		cfu/g	10	2202828	04/04/22	AOAC 991.14 (Petrifilm)		X
Total Coliforms	< LOQ		cfu/g	10	2202828	04/04/22	AOAC 991.14 (Petrifilm)		X
Mold (RAPID Petrifilm)	< LOQ		cfu/g	10	2202829	04/05/22	AOAC 2014.05 (RAPID)		X
Yeast (RAPID Petrifilm)	< LOQ		cfu/g	10	2202829	04/05/22	AOAC 2014.05 (RAPID)		X

Solvents Method Residual Solvents by GC/MS Units µg/g Batch 2202942 Analyze 04/06/22 08:55 AM

Analyte	Result	Limits	LOQ	Status	Notes	Analyte	Result	Limits	LOQ	Status	Notes
1,4-Dioxane	< LOQ	380	100	pass		2-Butanol	< LOQ	5000	200	pass	
2-Ethoxyethanol	< LOQ	160	30.0	pass		2-Methylbutane (Isopentane)	< LOQ		200		
2-Methylpentane	< LOQ		30.0			2-Propanol (IPA)	< LOQ	5000	200	pass	
2,2-Dimethylbutane	< LOQ		30.0			2,2-Dimethylpropane (neo-pentane)	< LOQ		200		
2,3-Dimethylbutane	< LOQ		30.0			3-Methylpentane	< LOQ		30.0		
Acetone	< LOQ	5000	200	pass		Acetonitrile	< LOQ	410	100	pass	
Benzene	< LOQ	2.00	1.00	pass		Butanes (sum)	< LOQ	5000	400	pass	
Cyclohexane	< LOQ	3880	200	pass		Ethyl acetate	< LOQ	5000	200	pass	
Ethyl benzene	< LOQ		200			Ethyl ether	< LOQ	5000	200	pass	
Ethylene glycol	< LOQ	620	200	pass		Ethylene oxide	< LOQ	50.0	20.0	pass	
Hexanes (sum)	< LOQ	290	150	pass		Isopropyl acetate	< LOQ	5000	200	pass	
Isopropylbenzene (Cumene)	< LOQ	70.0	30.0	pass		m,p-Xylene	< LOQ		200		
Methanol	< LOQ	3000	200	pass		Methylene chloride	< LOQ	600	60.0	pass	
Methylpropane (Isobutane)	< LOQ		200			n-Butane	< LOQ		200		
n-Heptane	< LOQ	5000	200	pass		n-Hexane	< LOQ		30.0		
n-Pentane	< LOQ		200			o-Xylene	< LOQ		200		
Pentanes (sum)	< LOQ	5000	600	pass		Propane	< LOQ	5000	200	pass	
Tetrahydrofuran	< LOQ	720	100	pass		Toluene	< LOQ	890	100	pass	
Total Xylenes	< LOQ		400			Total Xylenes and Ethyl benzene	< LOQ	2170	600	pass	

Pesticides Method AOAC 2007.01 & EN 15662 (mod) Units mg/kg Batch 2202980 Analyze 04/07/22 11:21 AM

Analyte	Result	Limits	Status	Notes
Multi-Residue Pesticide Profile ¹	< LOQ for all analytes			

Metals

Analyte	Result	Limits	Units	LOQ	Batch	Analyze	Method	Status	Notes
Arsenic	< LOQ	0.200	mg/kg	0.0137	2202970	04/06/22	AOAC 2013.06 (mod.)	pass	X
Cadmium	< LOQ	0.200	mg/kg	0.0137	2202970	04/06/22	AOAC 2013.06 (mod.)	pass	X
Lead	< LOQ	0.500	mg/kg	0.0137	2202970	04/06/22	AOAC 2013.06 (mod.)	pass	X
Mercury	< LOQ	0.100	mg/kg	0.00683	2202970	04/06/22	AOAC 2013.06 (mod.)	pass	X



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 22-003713/D002.R000
Report Date: 04/07/2022
ORELAP#: OR100028
Purchase Order:
Received: 04/01/22 11:10

Nutrition

Analyte	Result	Limits	Units	LOQ	Batch	Analyze	Method	Status	Notes
Moisture (Loss on Drying)	18.3		g/100g	0.10	2202929	04/04/22	AOAC 925.10 (mod.)		X
Water Activity	0.703		Aw	0.030	2202869	04/04/22	AOAC 978.18		X



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 22-003713/D002.R000
Report Date: 04/07/2022
ORELAP#: OR100028
Purchase Order:
Received: 04/01/22 11:10

These test results are representative of the individual sample selected and submitted by the client.

Abbreviations

Limits: Action Levels per OAR-333-007-0400, OAR-333-007-0210, OAR-333-007-0220, CCR title 16-division 42. BCC-section 5723

Limit(s) of Quantitation (LOQ): The minimum levels, concentrations, or quantities of a target variable (e.g., target analyte) that can be reported with a specified degree of confidence.

† = Analyte not NELAP accredited.

Units of Measure

cfu/g = Colony forming units per gram

g = g

g/100g = Grams per 100 Grams

µg/g = Microgram per gram

mg/kg = Milligram per kilogram = parts per million (ppm)

mg/4g = Milligram per 4g

% = Percentage of sample

Aw = Water Activity

% wt = µg/g divided by 10,000

Glossary of Qualifiers

X: Not ORELAP accredited.

Approved Signatory

Derrick Tanner
General Manager



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 22-003713/D002.R000
Report Date: 04/07/2022
ORELAP#: OR100028
Purchase Order:
Received: 04/01/22 11:10



Cannabis Multi-Residue Profile, Limits of Quantitation

Compound	LOQ (mg/kg)	Compound	LOQ (mg/kg)	Compound	LOQ (mg/kg)
Abamectin	0.100	Clethodim	0.050	Endrin	0.100
Acephate	0.100	Clethodim Sulfone	0.050	EPN	0.050
Acequinocyl	0.100	Clethodim Sulfoxide	0.050	EPTC	0.100
Acetamiprid	0.020	Clofentezine	0.020	Esfenvalerate/Fenvalerate	0.200
Acetochlor	0.100	Clomazone	0.020	Etaconazole	0.100
Acrinathrin	0.100	Clothianidin	0.200	Ethalfuralin	0.100
Alachlor	0.100	Coumaphos	0.050	Ethiofencarb	0.050
Aldicarb	0.100	Crotoxyphos	0.020	Ethion	0.200
Aldicarb sulfoxide	0.100	Cyanazine	0.020	Ethirimol	0.100
Aldoxycarb (Aldicarb-sulfone)	0.100	Cyanofenphos	0.020	Ethofumesate	0.050
Aldrin	0.100	Cyantranilprole	0.050	Ethoprophos	0.020
Ametocrtadin	0.020	Cyazofamid	0.020	Etofenprox	0.020
Ametryn	0.500	Cycloate	0.100	Etoxazole	0.020
Aspon	0.100	Cyfluthrin	0.200	Etridiazole	0.100
Asulam	0.100	Cyhalothrin, lambda	0.200	Etrimfos	0.020
Atrazine	0.100	Cymoxanil	0.050	Famoxadone	0.200
Atrazine-desethyl	0.100	Cypermethrin	0.200	Famphur	0.100
Azinphos-ethyl	0.020	Cyprodinil	0.100	Fenamidone	0.020
Azinphos-methyl	0.020	Dacthal	0.100	Fenamiphos	0.020
Azoxystrobin	0.020	Daminozide	0.100	Fenamiphos sulfone	0.020
Benalaxyl	0.020	DCPMU	0.050	Fenamiphos sulfoxide	0.020
Bendiocarb	0.020	DDD, o,p'-	0.100	Fenazaquin	0.100
Benfluralin	0.100	DDD, p,p'-	0.100	Fenbuconazole	0.100
Benoxacor	0.050	DDE, o,p'-	0.100	Fenchlorphos	0.100
Bensulide	0.050	DDE, p,p'-	0.100	Fenchlorphos-oxon	0.100
BHC alpha isomer	0.100	DDT, o,p'-	0.100	Fenhexamid	0.100
BHC beta isomer	0.100	DDT, p,p'-	0.100	Fenitrothion	0.100
BHC delta isomer	0.500	DEF (Tribufos)	0.100	Fenobucarb	0.050
Bifenazate	0.020	Deltamethrin	0.100	Fenoxycarb	0.020
Bifenthrin	0.020	Desmedipham	0.100	Fenpropathrin	0.050
Boscalid	0.020	Diallate	0.100	Fenpyroximate	0.020
Bromophos-ethyl	0.100	Diazinon	0.020	Fenson	0.100
Bromophos-methyl	0.200	Diazoxon	0.100	Fensulfothion	0.020
Bromopropylate	0.100	Dichlobenil	0.100	Fensulfothion oxon	0.020
Bromuconazole	0.100	Dichlofluanid	0.100	Fensulfothion sulfone	0.100
Bupirimate	0.020	Dichlorvos	0.100	Fensulfothion-oxon-sulfone	0.020
Buprofezin	0.050	Diclobutrazol	0.050	Fenthion	0.050
Butachlor	0.500	Dicofol	0.100	Fenthion oxon	0.020
Butralin	0.200	Dicrotophos	0.050	Fenthion oxon sulfone	0.100
Butylate	0.100	Dieldrin	0.100	Fenthion sulfone	0.050
Cadusafos	0.020	Diethofencarb	0.020	Fenuron	0.020
Captan	1.000	Diethyltoluamide (DEET)	0.050	Fipronil	0.100
Carbaryl	0.050	Difenoconazole	0.100	Flonicamid	0.100
Carbendazim	0.100	Dimethenamid	0.050	Fluchloralin	0.100
Carbofuran	0.020	Dimethoate	0.050	Flucythrinate	0.100
Carbophenothion	0.200	Dimethomorph	0.050	Fludioxonil	0.200
Carboxin	0.020	Diniconazole	0.200	Flufenacet	0.020
Carfentrazone-ethyl	0.100	Dinotefuran	0.200	Flumioxazin	0.100
Chlorantranilprole	0.020	Dioxathion	0.100	Fluometuron	0.020
Chlordane, cis-	0.200	Diphenamid	0.020	Fluopicolide	0.050
Chlordane, trans-	0.200	Diphenylamine	0.100	Fluopyram	0.020
Chlorfenapyr	0.500	Disulfoton	0.100	Fluoxastrobin	0.050
Chlorfenson	0.200	Disulfoton sulfone	0.100	Flupyradifurone	0.020
Chlorfenvinphos	0.050	Disulfoton sulfoxide	0.100	Fluridone	0.100
Chlorobenzilate	0.100	Diuron	0.050	Flusilazole	0.020
Chloroneb	0.200	Edifenphos	0.050	Flutolanil	0.020
Chlorpyrifos	0.050	Endosulfan alpha	0.200	Flutriafol	0.020
Chlorpyrifos-methyl	0.200	Endosulfan beta	0.200	Fluxalinat, tau-	0.100
CIPC	1.000	Endosulfan sulfate	0.100	Fluxapyroxad	0.020



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 22-003713/D002.R000
Report Date: 04/07/2022
ORELAP#: OR100028
Purchase Order:
Received: 04/01/22 11:10



Cannabis Multi-Residue Profile, Limits of Quantitation

Compound	LOQ (mg/kg)	Compound	LOQ (mg/kg)	Compound	LOQ (mg/kg)
Fomesafen	0.100	Mexacarbate	0.020	Propamocarb	0.050
Fonofos	0.100	MGK 264	0.020	Propanil	0.050
Forchlorfenuron	0.050	Mirex	0.100	Propargite	0.050
Formetanate	0.050	Molinate	0.050	Propazine	0.020
Furathiocarb	0.020	Monocrotophos	0.100	Propetamphos	0.050
Heptachlor	0.100	Monolinuron	0.020	Propham	0.050
Heptachlor epoxide	0.100	Myclobutanil	0.050	Propiconazole	0.050
Heptenophos	0.100	Naled	0.100	Propoxur	0.050
Hexachlorobenzene	0.100	Napropamide	0.050	Propoxycarbazone Na	0.050
Hexaconazole	0.100	Neburon	0.020	Propyzamide	0.050
Hexazinone	0.100	Nitrapyrin	0.100	Prothiofos	0.100
Hexythiazox	0.020	Norflurazon	0.050	Pyraclostrobin	0.020
Imazalil	0.100	Omethoate	0.100	Pyrazophos	0.050
Imidacloprid	0.100	O-Phenylphenol	0.100	Pyrethrins	0.050
Indaziflam	0.020	Oxadixyl	0.100	Pyridaben	0.020
Indoxacarb	0.020	Oxamyl	0.100	Pyridafol	0.100
Iprobenfos	0.100	Oxamyl-oxime	0.100	Pyridate	0.020
Iprodione	0.100	Oxychlorane	0.100	Pyrimethanil	0.050
Isobenzan	0.100	Oxydemeton-Methyl	0.100	Pyriproxifen	0.020
Isocarbophos	0.500	Oxythioquinox	0.200	Pyroxasulfone	0.020
Isodrin	0.100	Paclobutrazol	0.050	Pyroxulam	0.020
Isofenphos	0.050	Paraoxon-ethyl	0.020	Quinalphos	0.050
Isofenphos-methyl	0.020	Paraoxon methyl	0.100	Quinoxyfen	0.050
Isofenphos oxon	0.050	Parathion ethyl	0.100	Quintozene (PCNB)	0.200
Isoprocarb	0.020	Parathion methyl	0.200	Resmethrin	0.050
Isopropalin	0.200	Penconazole	0.050	Rotenone	0.050
Isoprothiolane	0.050	Pendimethalin	0.050	S421	0.100
Isoproturon	0.050	Penflufen	0.020	Simazine	0.100
Isoxaben	0.050	Pentachloroaniline	0.100	Simetryn	0.200
Isoxaflutole	0.050	Pentachloroanisole	0.100	Spinetoram	0.020
Kresoxim-methyl	0.050	Pentachlorobenzene (PCB)	0.100	Spinosad	0.050
Lactofen	0.500	Pentachlorothioanisole (PCTA)	0.100	Spirodiclofen	0.100
Lenacil	0.100	Penthiopyrad	0.020	Spiromesifen	0.050
Lindane (gamma BHC)	0.100	Permethrin	0.050	Spirotetramat	0.050
Linuron	0.020	Perthane	0.100	Spiroxamine	0.020
Malaaxon	0.050	Phenmedipham	0.050	Sulfotep	0.050
Malathion	0.050	Phenthoate	0.050	Sulfoxaflor	0.050
Mandipropamid	0.020	Phorate	0.050	Sulprofos	0.020
Mecarbam	0.020	Phorate Sulfone	0.050	Tebuconazole	0.100
Mepanipyrim	0.050	Phorate Sulfoxide	0.050	Tebufenozide	0.020
Merphos	0.500	Phosalone	0.050	Tebuthiuron	0.020
Metalaxyl	0.050	Phosmet	0.100	Tecnazene	0.100
Metaldehyde	0.050	Phosphamidon	0.050	Tefluthrin	0.100
Metconazole	0.100	Phoxim	0.050	Terbufos	0.020
Methacrifos	0.100	Pinoxaden	0.020	Terbufos sulfone	0.050
Methamidophos	0.050	Piperonyl butoxide	0.050	Terbufos sulfoxide	0.050
Methidathion	0.050	Pirimicarb	0.020	Terbutylazine	0.020
Methiocarb	0.050	Pirimiphos-methyl	0.050	Terbutryn	0.020
Methiocarb sulfone	0.100	Pirimiphos-ethyl	0.020	Tetrachlorvinphos	0.050
Methiocarb sulfoxide	0.100	Prallethrin	0.100	Tetraconazole	0.050
Methomyl	0.100	Prochloraz	0.020	Tetradifon	0.200
Methoxychlor	0.100	Procymidone	0.100	Tetramethrin	0.050
Methoxyfenozide	0.020	Profenofos	0.100	Tetrasul	0.100
Metobromuron	0.050	Profluralin	0.100	Thiabendazole	0.100
Metolachlor	0.100	Promecarb	0.050	Thiabendazole, 5-hydroxy	0.100
Metolcarb	0.050	Prometon	0.100	Thiacloprid	0.050
Metrafenone	0.050	Prometryn	0.020	Thiamethoxam	0.100
Metribuzin	0.100	Propachlor	0.020	Thiobencarb	0.050
Mevinphos	0.100			Thiodicarb	0.050
				Thiophanate-methyl	0.050



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 22-003713/D002.R000
Report Date: 04/07/2022
ORELAP#: OR100028
Purchase Order:
Received: 04/01/22 11:10



Cannabis Multi-Residue Profile, Limits of Quantitation

Compound	LOQ (mg/kg)	Compound	LOQ (mg/kg)	Compound	LOQ (mg/kg)
Tolclofos-methyl	0.100	Triazophos	0.020	Trifloxystrobin	0.020
Triforin	0.100	Tolyfluanid	0.050	Triticonazole	0.050
Tralkoxydim	0.100	Tridiphane	0.500	Vinclozolin	0.100
Triadimefon	0.050	Triflumizole	0.020	Zoxamide	0.020
Triallate	0.100	Trifluralin	0.100		

LOQ = Limit of Quantitation, mg/kg

Factors affecting the LOQ include instrumentation sensitivity for a particular analyte, sample size, moisture content (percent solids) of the sample, effectiveness of the cleanup on the sample extract, and especially the type of sample matrix.



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 22-003713/D002.R000
Report Date: 04/07/2022
ORELAP#: OR100028
Purchase Order:
Received: 04/01/22 11:10

3713



12423 NE Whitaker Way Portland OR 97230 p.503-254-1794

Cannabis Chain of Custody Record

ORELAP ID: OR100028

Company: NW Natural Goods		Analysis Requested											Purchase Order Number:						
Contact: Annie Nair		Pesticides - OR 59 compounds	Pesticide Multi-Residue - 379 compounds	Potency	Residual Solvents	Water Activity	Moisture	Terpenes	Micro: Yeast and Mold	Micro: E. Coli and Total Coliform	Heavy Metals	Mycotoxins	Other	OPEN MARKET			Project Number:		
Address: 11791 SE HWY 212, Clackamas, OR 97015														C L P U			Project Name:		
Email: annienair@nwnaturalgoods.com																	<input type="checkbox"/> Report Instructions <input type="checkbox"/> Send to State - METRC <input checked="" type="checkbox"/> Email Final Results: <input type="checkbox"/> Fax Final Results <input type="checkbox"/> Cash/Check/CC/Net 30		
Phone: 503-453-4219 Fax:																	Other:		
Processor's License: 330-1058115IHH																	Matrix	Weight	Serving size for edibles
Field ID	Date/Time Collected																		
HEMP - BB 0071	4/1/22	X	X	X	X	X	X	X	X	X				edible	80g	4g			

Collected By:	Relinquished By:	Date	Time	Received by:	Date	Time	Lab Use Only:
<input checked="" type="checkbox"/> Standard (5 day)	Annie Nair	4/1/22	1030	[Signature]	4.1.22	1050	Client Alias:
<input type="checkbox"/> Rush (3-4 day) (1.5x Standard)	[Signature]	4.1.22	1110	[Signature]	4/1/22	1116	Order Number:
<input type="checkbox"/> Priority Rush (2 day) (2x Standard)							Proper Container
							Sample Condition
							Temperature: 21.6
							Shipped Via:
							Evidence of cooling: <input type="checkbox"/> Yes <input type="checkbox"/> No

SUBMISSION OF SAMPLES WITH TESTING REQUIREMENTS TO PIXIS WILL BE UNDERSTOOD TO BE AN AGREEMENT FOR SERVICES IN ACCORDANCE WITH THE CONDITIONS LISTED ON THE BACK OF THIS FORM
 Revision: 1.00 Control#: CF023 www.pixislabs.com
 Effective 11/8/2018 Revised 11/8/2018 Page 1 of 2



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 22-003713/D002.R000
Report Date: 04/07/2022
ORELAP#: OR100028
Purchase Order:
Received: 04/01/22 11:10

Revision: 1 Document ID: 7148
Legacy ID: Worksheet Validated 04/20/2021

Laboratory Quality Control Results

J AOAC 2015 V98-6 **Batch ID: 2202906**

Laboratory Control Sample									
Analyte	Result	Spike	Units	% Rec	Limits		Evaluation	Notes	
CBDVA	0.0357	0.033	%	107	80.0	- 120	Acceptable		
CBDV	0.0396	0.033	%	119	80.0	- 120	Acceptable		
CBE	0.0344	0.033	%	103	80.0	- 120	Acceptable		
CBDA	0.0351	0.035	%	98.8	90.0	- 110	Acceptable		
CBGA	0.0350	0.033	%	105	80.0	- 120	Acceptable		
CBG	0.0358	0.033	%	107	80.0	- 120	Acceptable		
CBD	0.0348	0.035	%	100	90.0	- 110	Acceptable		
THCV	0.0366	0.033	%	110	80.0	- 120	Acceptable		
d8THCV	0.0365	0.033	%	110	80.0	- 120	Acceptable		
THCVA	0.0337	0.033	%	101	80.0	- 120	Acceptable		
CBN	0.0367	0.036	%	102	90.0	- 110	Acceptable		
exo-THC	0.0325	0.033	%	97.5	80.0	- 120	Acceptable		
d9THC	0.0361	0.036	%	101	90.0	- 110	Acceptable		
d8THC	0.0355	0.033	%	106	90.0	- 110	Acceptable		
CBL	0.0336	0.033	%	101	80.0	- 120	Acceptable		
CBC	0.0365	0.033	%	110	80.0	- 120	Acceptable		
THCA	0.0340	0.034	%	99.2	90.0	- 110	Acceptable		
CBCA	0.0350	0.033	%	105	80.0	- 120	Acceptable		
CBLA	0.0323	0.033	%	97.0	80.0	- 120	Acceptable		
CBT	0.0332	0.033	%	99.7	80.0	- 120	Acceptable		

Method Blank

Analyte	Result	LOQ	Units	Limits		Evaluation	Notes	
CBDVA	< LOQ	0.005	%	< 0.005		Acceptable		
CBDV	< LOQ	0.005	%	< 0.005		Acceptable		
CBE	< LOQ	0.005	%	< 0.005		Acceptable		
CBDA	< LOQ	0.005	%	< 0.005		Acceptable		
CBGA	< LOQ	0.005	%	< 0.005		Acceptable		
CBG	< LOQ	0.005	%	< 0.005		Acceptable		
CBD	< LOQ	0.005	%	< 0.005		Acceptable		
THCV	< LOQ	0.005	%	< 0.005		Acceptable		
d8THCV	< LOQ	0.005	%	< 0.005		Acceptable		
THCVA	< LOQ	0.005	%	< 0.005		Acceptable		
CBN	< LOQ	0.005	%	< 0.005		Acceptable		
exo-THC	< LOQ	0.005	%	< 0.005		Acceptable		
d9THC	< LOQ	0.005	%	< 0.005		Acceptable		
d8THC	< LOQ	0.005	%	< 0.005		Acceptable		
CBL	< LOQ	0.005	%	< 0.005		Acceptable		
CBC	< LOQ	0.005	%	< 0.005		Acceptable		
THCA	< LOQ	0.005	%	< 0.005		Acceptable		
CBCA	< LOQ	0.005	%	< 0.005		Acceptable		
CBLA	< LOQ	0.005	%	< 0.005		Acceptable		
CBT	< LOQ	0.005	%	< 0.005		Acceptable		

Abbreviations

ND - None Detected at or above MRL
RPD - Relative Percent Difference
LOQ - Limit of Quantitation

Units of Measure:

% - Percent



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 22-003713/D002.R000
Report Date: 04/07/2022
ORELAP#: OR100028
Purchase Order:
Received: 04/01/22 11:10

Revision: 1 Document ID: 7148
 Legacy ID: Worksheet Validated 04/20/2021

Laboratory Quality Control Results

J AOAC 2015 V98-6		Batch ID: 2202906						
Sample Duplicate		Sample ID: 22-001950-0003						
Analyte	Result	Org. Result	LOQ	Units	RPD	Limits	Evaluation	Notes
CBDVA	< LOQ	< LOQ	0.003	%	NA	< 20	Acceptable	
CBDV	< LOQ	< LOQ	0.003	%	NA	< 20	Acceptable	
CBE	< LOQ	< LOQ	0.003	%	NA	< 20	Acceptable	
CBDA	< LOQ	< LOQ	0.003	%	NA	< 20	Acceptable	
CBGA	< LOQ	< LOQ	0.003	%	NA	< 20	Acceptable	
CBG	0.00376	0.00370	0.003	%	1.66	< 20	Acceptable	
CBD	< LOQ	< LOQ	0.003	%	NA	< 20	Acceptable	
THCV	< LOQ	< LOQ	0.003	%	NA	< 20	Acceptable	
d8THCV	< LOQ	< LOQ	0.003	%	NA	< 20	Acceptable	
THCVA	< LOQ	< LOQ	0.003	%	NA	< 20	Acceptable	
CBN	0.130	0.128	0.003	%	1.65	< 20	Acceptable	
exo-THC	< LOQ	< LOQ	0.003	%	NA	< 20	Acceptable	
d9THC	0.128	0.126	0.003	%	1.89	< 20	Acceptable	
d8THC	< LOQ	< LOQ	0.003	%	NA	< 20	Acceptable	
CBL	< LOQ	< LOQ	0.003	%	NA	< 20	Acceptable	
CBC	< LOQ	< LOQ	0.003	%	NA	< 20	Acceptable	
THCA	< LOQ	< LOQ	0.003	%	NA	< 20	Acceptable	
BCA	< LOQ	< LOQ	0.003	%	NA	< 20	Acceptable	
CBLA	< LOQ	< LOQ	0.003	%	NA	< 20	Acceptable	
CBT	< LOQ	< LOQ	0.003	%	NA	< 20	Acceptable	

Abbreviations

- ND - None Detected at or above MRL
- RPD - Relative Percent Difference
- LOQ - Limit of Quantitation

Units of Measure:

% - Percent



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 22-003713/D002.R000
Report Date: 04/07/2022
ORELAP#: OR100028
Purchase Order:
Received: 04/01/22 11:10

Revision: Document ID:
 Legacy ID: Effective:

Laboratory Quality Control Results									
Residual Solvents									
Batch ID: 2202942									
Method Blank				Laboratory Control Sample					
Analyte	Result	LOQ	Notes	Result	Spike	Units	% Rec	Limits	Notes
Propane	ND	< 200		535	572	µg/g	93.5	60 - 120	
Isobutane	ND	< 200		749	731	µg/g	102.5	60 - 120	
Butane	ND	< 200		766	731	µg/g	104.8	60 - 120	
2,2-Dimethylpropane	ND	< 200		897	936	µg/g	95.8	60 - 120	
Methanol	ND	< 200		1540	1620	µg/g	95.1	60 - 120	
Ethylene Oxide	ND	< 30		59.5	56.2	µg/g	105.9	60 - 120	
2-Methylbutane	ND	< 200		1590	1620	µg/g	98.1	60 - 120	
Pentane	ND	< 200		1540	1610	µg/g	95.7	60 - 120	
Ethanol	ND	< 200		1590	1630	µg/g	97.5	70 - 130	
Ethyl Ether	ND	< 200		1480	1620	µg/g	91.4	60 - 120	
2,2-Dimethylbutane	ND	< 30		157	174	µg/g	90.2	60 - 120	
Acetone	ND	< 200		1360	1690	µg/g	82.4	60 - 120	
2-Propanol	ND	< 200		1430	1610	µg/g	88.8	60 - 120	
Ethyl Formate	ND	< 500		1410	1600	µg/g	88.1	70 - 130	
Acetonitrile	ND	< 100		472	496	µg/g	94.8	60 - 120	
Methyl Acetate	ND	< 500		1490	1610	µg/g	92.5	70 - 130	
2,3-Dimethylbutane	ND	< 30		161	176	µg/g	91.5	60 - 120	
Dichloromethane	ND	< 60		486	510	µg/g	95.3	60 - 120	
2-Methylpentane	ND	< 30		150	176	µg/g	85.2	60 - 120	
MTBE	ND	< 500		1440	1600	µg/g	90.0	70 - 130	
3-Methylpentane	ND	< 30		155	175	µg/g	88.6	60 - 120	
Hexane	ND	< 30		173	177	µg/g	97.7	60 - 120	
1-Propanol	ND	< 500		1510	1610	µg/g	93.8	70 - 130	
Methyl ethyl ketone	ND	< 500		1510	1600	µg/g	94.4	70 - 130	
Ethyl acetate	ND	< 200		1500	1630	µg/g	92.0	60 - 120	
2-Butanol	ND	< 200		1430	1620	µg/g	88.3	60 - 120	
Tetrahydrofuran	ND	< 100		437	500	µg/g	87.4	60 - 120	
Cyclohexane	ND	< 200		1390	1620	µg/g	85.8	60 - 120	
2-methyl-1-propanol	ND	< 500		1310	1620	µg/g	80.9	70 - 130	
Benzene	ND	< 1		4.54	5.32	µg/g	85.3	60 - 120	
Isopropyl Acetate	ND	< 200		1370	1620	µg/g	84.6	60 - 120	
Heptane	ND	< 200		1400	1770	µg/g	79.1	60 - 120	
1-Butanol	ND	< 500		1250	1600	µg/g	78.1	70 - 130	
Propyl Acetate	ND	< 500		1450	1600	µg/g	90.6	70 - 130	
1,4-Dioxane	ND	< 100		423	504	µg/g	83.9	60 - 120	
2-Ethoxyethanol	ND	< 30		156	181	µg/g	86.2	60 - 120	
Methylisobutylketone	ND	< 500		1360	1610	µg/g	84.5	70 - 130	
3-Methyl-1-butanol	ND	< 500		1200	1610	µg/g	74.5	70 - 130	
Ethylene Glycol	ND	< 200		361	494	µg/g	73.1	60 - 120	
Toluene	ND	< 200		430	491	µg/g	87.6	60 - 120	
Isobutyl Acetate	ND	< 500		1500	1600	µg/g	93.8	70 - 130	
1-Pentanol	ND	< 500		1530	1610	µg/g	95.0	70 - 130	
Butyl Acetate	ND	< 500		1470	1610	µg/g	91.3	70 - 130	
Ethylbenzene	ND	< 200		861	973	µg/g	88.5	60 - 120	
m,p-Xylene	ND	< 200		857	996	µg/g	86.0	60 - 120	
o-Xylene	ND	< 200		890	973	µg/g	91.5	60 - 120	
Cumene	ND	< 30		141	170	µg/g	82.9	60 - 120	
Anisole	ND	< 500		1350	1610	µg/g	83.9	70 - 130	
DMSO	ND	< 500		1560	1630	µg/g	95.7	70 - 130	
1,2-dimethoxyethane	ND	< 50		148	164	µg/g	90.2	70 - 130	
Triethylamine	ND	< 500		1360	1600	µg/g	85.0	70 - 130	
N,N-dimethylformamide	ND	< 150		430	497	µg/g	86.5	70 - 130	
N,N-dimethylacetamide	ND	< 150		488	498	µg/g	98.0	70 - 130	
Pyridine	ND	< 50		175	180	µg/g	97.2	70 - 130	
1,2-Dichloroethane	ND	< 1		1.01	1	µg/g	101.0	70 - 130	
Chloroform	ND	< 1		1.03	1	µg/g	103.0	70 - 130	
Trichloroethylene	ND	< 1		1.14	1	µg/g	114.0	70 - 130	



Revision: Document ID:
Legacy ID: Effective:

QC - Sample Duplicate Sample ID: 22-003693-0004

Analyte	Result	Org. Result	LOQ	Units	RPD	Limits	Accept/Fail	Notes
Propane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Isobutane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Butane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2,2-Dimethylpropane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Methanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethylene Dioxide	ND	ND	30	µg/g	0.0	< 20	Acceptable	
2-Methylbutane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Pentane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethyl Ether	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2,2-Dimethylbutane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Acetone	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2-Propanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethyl Formate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Acetonitrile	ND	ND	100	µg/g	0.0	< 20	Acceptable	
Methyl Acetate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
2,3-Dimethylbutane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Dichloromethane	ND	ND	60	µg/g	0.0	< 20	Acceptable	
2-Methylpentane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
MIBK	ND	ND	500	µg/g	0.0	< 20	Acceptable	
1,3-Methylpentane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Hexane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
1-Propanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Methyl ethyl ketone	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Ethyl acetate	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2-Butanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Tetrahydrofuran	ND	ND	100	µg/g	0.0	< 20	Acceptable	
Cyclohexane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2-methyl-1-propanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Benzene	ND	ND	1	µg/g	0.0	< 20	Acceptable	
Isopropyl Acetate	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Heptane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
1-Butanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Propyl Acetate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
1,4-Dioxane	ND	ND	100	µg/g	0.0	< 20	Acceptable	
2-Ethoxyethanol	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Methylisobutylketone	ND	ND	500	µg/g	0.0	< 20	Acceptable	
3-Methyl-1-butanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Ethylene Glycol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Toluene	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Isobutyl Acetate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
1-Pentanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Butyl Acetate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Ethylbenzene	ND	ND	200	µg/g	0.0	< 20	Acceptable	
m,p-Xylene	ND	ND	200	µg/g	0.0	< 20	Acceptable	
o-Xylene	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Cumene	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Anisole	ND	ND	500	µg/g	0.0	< 20	Acceptable	
DMSO	ND	ND	500	µg/g	0.0	< 20	Acceptable	
1,2-dimethoxyethane	ND	ND	50	µg/g	0.0	< 20	Acceptable	
Triethylamine	ND	ND	500	µg/g	0.0	< 20	Acceptable	
N,N-dimethylformamide	ND	ND	150	µg/g	0.0	< 20	Acceptable	
N,N-dimethylacetamide	ND	ND	150	µg/g	0.0	< 20	Acceptable	
Pyridine	ND	ND	50	µg/g	0.0	< 20	Acceptable	
1,2-Dichloroethane	ND	ND	1	µg/g	0.0	< 20	Acceptable	
Chloroform	ND	ND	1	µg/g	0.0	< 20	Acceptable	
Trichloroethylene	ND	ND	1	µg/g	0.0	< 20	Acceptable	

Abbreviations

ND - None Detected at or above MRL
RPD - Relative Percent Difference
LOQ - Limit of Quantitation

Units of Measure:

µg/g - Microgram per gram or ppm



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 22-003713/D002.R000
Report Date: 04/07/2022
ORELAP#: OR100028
Purchase Order:
Received: 04/01/22 11:10





12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 22-003713/D002.R000
Report Date: 04/07/2022
ORELAP#: OR100028
Purchase Order:
Received: 04/01/22 11:10

Explanation of QC Flag Comments:

Code	Explanation
Q	Matrix interferences affecting spike or surrogate recoveries.
Q1	Quality control result biased high. Only non-detect samples reported.
Q2	Quality control outside QC limits. Data considered estimate.
Q3	Sample concentration greater than four times the amount spiked.
Q4	Non-homogenous sample matrix, affecting RPD result and/or % recoveries.
Q5	Spike results above calibration curve.
Q6	Quality control outside QC limits. Data acceptable based on remaining QC.
R	Relative percent difference (RPD) outside control limit.
R1	RPD non-calculable, as sample or duplicate results are less than five times the LOQ.
R2	Sample replicates RPD non-calculable, as only one replicate is within the analytical range.
LOQ1	Quantitation level raised due to low sample volume and/or dilution.
LOQ2	Quantitation level raised due to matrix interference.
B	Analyte detected in method blank, but not in associated samples.
B1	The sample concentration is greater than 5 times the blank concentration.
B2	The sample concentration is less than 5 times the blank concentration.